

FPL Turkey Point Monitoring Plan
Field Instrument Calibration Form

Parameter: **Temperature, water**

FDEP-SOP Reference: **FT1400**

QAPP Requirements: Monthly verification against NIST-traceable thermometer
Must be within $\pm 0.5^{\circ}\text{C}$ of NIST-traceable readings
Quarterly verification at temperatures above and below the range
of sample readings for the quarter

Date	Time	Instrument or meter ID	NIST- Traceable Thermometer ID	Instrument or meter reading (Deg C)	NIST Thermometer reading (Deg C)	Acceptance Criteria Met? (Y/N)	Comments ¹	Calibration verified by

¹ = Indicate any failed verifications; all corrective actions taken; any maintenance performed.